

**Corrigendum-1****Notice: GEM/2025/B/6332908**

ઇન્સ્ટીટ્યુટ ઓફ સિસ્મોલોજીકલ રિસર્ચ, નોલેઝ કોરિડોર,  
રાયસણ, ગાંધીનગર-૩૮૨૦૦૯, ફોન.૦૭૯-૬૬૭૩૯૦૦૯/૦૭

**સ્પેશીફિકેશન અંગેના સુધારા બાબત**

ઇન્સ્ટીટ્યુટ ઓફ સિસ્મોલોજીકલ રિસર્ચના ટેન્ડર નોટીસ નંબર:  
- **GEM/2025/B/6332908**, [www.isr.gujarat.gov.in](http://www.isr.gujarat.gov.in) ની વેબસાઇટ ઉપર પ્રસિધ્ધ કરવામાં આવેલ છે. જેની શરતોમાં નીચે મુજબ સુધારો કરેલ છે.

<b>Our specification</b>	<b>Vendor's (M/s Shree siddhivinayak enterprise) query/submission</b>	<b>Reply/Revised Specification</b>
Object field: should include 29mm to 1.5mm or better	Should include 31 mm - 1.5mm or better	Object field: should include 31mm to 1.5mm or better
Observation tube: Ergonomic tube with 5-30 viewing angle	Trinocular tube of 20° angle	Observation tube: Trinocular tube with 20°(±10°) or better
Camera and software: Exposure time 0.02msec to 15 Sec or better variable color depth 4,8 and 24bit Z-capturing with EDF - Extended depth of Field	Exposure time: 100 microsecond 4 s Remove As stand is not motorized, EDF Will be manual.	Camera and software: Exposure time 100msec to 4 Sec or better. Extended depth of Field is needed.
Accessories	You can add a printer to print the report or the captured photographs.	Should provide B/w LaserJet Printer with duplex facility, WiFi (make: HP, Canon, Brother)

તા: ૨૦/૦૬/૨૦૨૫

સ્ટોર અને પરચેઝ અધિકારી

ઇન્સ્ટીટ્યુટ ઓફ સિસ્મોલોજીકલ રિસર્ચ

